

6 EASY STEPS TO ATTACH VELCRO TO ORFICAST

NO SEWING, GLUING OR IRONING! ONLY DRY HEAT.

1. Prepare two strips of Orficast 1" x 2", a small piece of adhesive backed hook and a longer piece of loop strapping.

- 2. Using dry heat, heat up one strip of Orficast and the adhesive side of the hook and press together. The Orficast acts as an interface
- 3. Trim the excess Orficast and adhere to the orthosis using dry heat on both the orthosis and the prepared hook.
- 4. Heat the second strip of Orficast and wrap around one end of the loop strapping. Trim corners.
- 5. Heat the orthosis and the prepared end of the loop strap with dry heat and press together.
- 6. Alternatively, heat a thin strip of Orficast and use as binding tape around the edges of the hook and the loop after these have been bonded to the orthosis.











ORFIT INDUSTRIES

Vosveld 9A | B-2110 Wijnegem | Belgium T (+32) (0)3 326 20 26 welcome@orfit.com

ORFIT INDUSTRIES AMERICA

350 Jericho Turnpike, Suite 302 | Jericho, New York 11753 T 516-935-8500 | F 516-935-8505 sales-us@orfit.com